

## 14. HPLC certified Plastic Disposable Syringes with Luer Lock and Luer Slip

HPLC certified, non sterile disposable syringes made of solvent robust polypropylene for all syringe filter applications

- Each manufactured batch is HPLC controlled and supplied with an appropriate certificate upon request.
- Syringes manufactured according to ISO 13485 (Medical devices).
- Luer Lock manufactured according to ISO 594-2/DIN EN 1707.
- Luer Slip manufactured according to ISO 594-1/DIN EN 20594-1.
- Free of latex, free of plasticizers, free of PVC.
- Two-part, all-plastic construction made of a chemically resistant, inert polypropylene.
- No rubber plunger seals or silicone lubricants that may cause sample contamination.
- Safe functional back-stop feature.
- Easy to read permanent graduations.
- All Luer Lock syringes have centered tips.
- Available with Luer Lock and Luer Slip connections.



### 14.1 Plastic Disposable Syringes with Luer Lock

Scale 1:2



Art. No.	02 36 2399	05 36 2400	10 36 2401	20 36 2402
Description	2ml Disposable Syringe non sterile <b>Luer Lock</b>	5ml Disposable Syringe non sterile <b>Luer Lock</b>	10ml Disposable Syringe non sterile <b>Luer Lock</b>	20ml Disposable Syringe non sterile <b>Luer Lock</b>

100 pcs. per PE-Bag

### 14.2 Plastic Disposable Syringes with Luer Slip

Scale 1:2



Art. No.	02 36 2403	05 36 2404	10 36 2405	20 36 2406
Description	2ml Disposable Syringe non sterile <b>Luer Slip</b>	5ml Disposable Syringe non sterile <b>Luer Slip</b>	10ml Disposable Syringe non sterile <b>Luer Slip</b>	20ml Disposable Syringe non sterile <b>Luer Slip</b>

100 pcs. per PE-Bag