

ANKORIS

MONOFOCAL OPTIC



PhysIOL

ADVANCED OPTICAL SOLUTIONS



ANKORIS

Monofocal toric IOL

Technical specifications

Technical name	Ankoris						
Material	26% hydrophilic acrylic						
Overall diameter	11.40 mm						
Optic diameter	6.00 mm						
Optic	Biconvex aspheric aberration-correcting (-0.11 μ SA)						
Filtration	UV and blue light						
Refractive index	1.46						
Abbe number	58						
Angulation	5°						
Injection system	Medicel Accuject 2.0 from 6D to 24.5D Medicel Accuject 2.1 or 2.2 from 25D to 30D						
Incision size	≥ 2.0 mm						
Spherical power	6D to 30D (0.5D steps)						
Cylinder power (IOL plane)	6D to 9.5D spherical power: 1.50 - 2.25 - 3.00 - 3.75D (on demand: 4.50 - 5.25 - 6.00D) 10D to 30D spherical power: 1.50 - 2.25 - 3.00 - 3.75 - 4.50 - 5.25 - 6.00D						
Square edge	360°						
Nominal manufacturer A constant	118.50						
Suggested A constant*			Interferometry			Ultrasound	
	Hoffer Q: pACD		5.59			5.35	
	Holladay 1: Sf		1.83			1.57	
	SRK II: A		119.31			119.06	
	SRK/T: A		118.95			118.73	
Haigis**: a0; a1; a2		1.36; 0.4; 0.1			1.13; 0.4; 0.1		
Technical name	Ankoris 1.5	Ankoris 2.25	Ankoris 3.0	Ankoris 3.75	Ankoris 4.5	Ankoris 5.25	Ankoris 6.0
Cylinder power at IOL plane	1.50D	2.25D	3.00D	3.75D	4.50D	5.25D	6.00D
Cylinder power at corneal plane	1.03D	1.55D	2.06D	2.57D	3.08D	3.60D	4.11D
Recommended corneal astigmatism correction range	0.90D - 1.28D	1.29D - 1.80D	1.81D - 2.32D	2.33D - 2.82D	2.83D - 3.33D	3.34D - 3.85D	3.86D - 4.36D

* Estimates only; surgeons are recommended to use their own values based upon their personal experience. Refer to our website for updates.

** Not optimized.

INJECTION GUIDELINES






The Medice Accuject injection system is recommended for implanting the Ankoris lenses.

This fully single-use system represents total reliability for safe and effective lens injections.

Its compact design with integrated cartridge enables a simple, predictable loading and positioning of the lens.

Accuject 2.0 for lens diopters < 25D

Accuject 2.1 or 2.2 for lens diopters ≥ 25D



1. Apply viscoelastic into the tip and the loading chamber of the injector cartridge.
2. Remove the lens from the lens holder. Position the lens into the cartridge in such way that the two haptics with the holes are pointing at 1 and 7 o'clock.
3. Exert slight pressure onto the lens optic and make sure that all haptics are inside before further closing the cartridge. Close the cartridge and check the position of the lens.
4. Once the "click-lock" mechanism engages, the lens is securely loaded and ready for injection.
5. Press the injector plunger forward and push the lens into the conical tip of the cartridge.
6. Pull the plunger back a few millimeters and then inject the lens in one continuous motion. For gently implantation, it is not necessary to push the plunger until the end of the cartridge.

SURGICAL GUIDELINES

Preoperative:

1. Use the PhysiOL toric calculator www.physioltoric.eu which will recommend you the spherical and cylindrical lens powers and the optimal axis alignment of the IOL.
2. Mark the eye with the patient sitting upright in order to avoid cyclotorsion effect.

Peroperative:

1. When the Ankoris lens is injected in the capsular bag, remove all viscoelastic behind and in front of the lens using I/A canula.
2. With a syringe filled with BSS solution, test the watertight self-sealing of the incisions and ensure that the normal intraocular pressure is recovered.
3. If necessary, reposition the lens in the axis of the IOL marks using a micromanipulator.
4. Gently push the lens towards the posterior capsule with the micromanipulator.
5. Check again that the incision is watertight.
6. Carefully remove the eyelid speculum.

Do not over-inflate the capsular bag at the end of the surgery.

Distributed by

