

## AUTOMATED DIGITAL THICKNESS GAUGE FOR DETERMINATION OF CONDOM AND GLOVE THICKNESS

A foot pedal operated tester with digital indicator and data connection. The operator can use both hands to handle the sample to be tested.

The indicator can be adjusted in various positions as preferred by the operator. The indicator is provided with a data output, adapters and software for connection to a PC for importing test data.

The measuring force is by delivery adjusted to the requirements for measuring of condoms and gloves (latex film).

Supplied with a traceable calibration certificate.

### Specifications

Measuring range:	0-12mm
Resolution:	0,001mm
Measuring accuracy:	±0,003mm
Contact point:	Ø 6,3mm flat face carbide
Parallelism between spindle and measuring surfaces:	3µm
Pressure on the foot:	22±4 kPa

**Art. No.: 402**



*VALENDOR reserves the right to modify this specification in part or as a whole*